

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Lab Name Govt. PHC, Konalur Lab No 12924

Constituent Group Chemistry III Date of Result Entered : 20/07/2023

PT item Lyophilized human serum based Date of Report Published : 07/08/2023

| SI.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your | SDI | - 11 |
|-------|----------------|--|--|-----------------------|--------|--------------|------|--------------|-------|------|
| | | | | | | CV | SD | Value | SDI | U |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1876 | 109.35 | 5.55 | 6.07 | 117 mg/dL | 1.26 | 0.28 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1850 | 5.83 | 6.00 | 0.35 | 6 mg/dL | 0.49 | 0.02 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1776 | 113.00 | 4.12 | 4.66 | 110 mg/dL | -0.64 | 0.22 |

| SDI Range | Interpretation | | | |
|-----------------------|--|--|--|--|
| Within -1.00 to +1.00 | Excellent. | | | |
| Within ±1.01 to ±2.00 | Good. | | | |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. | | | |
| Beyond ±3.0 | Unacceptable performance. Action Signal. | | | |

LAB ADDRESS:

Govt. PHC, Konalur

Konalur TAMILNADU

Panela Christudoss

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******