

**VIEW LAB MONTHLY SUMMARY**

**Lab Name** Govt. PHC, Periakayapakkam  
**Lab No** 12242  
**Month** March  
**Year** 2023  
**Constituent Group** Chemistry III

**Details About Robust Analysis**  
**Detail About Monthly Summary**  
**Detail about SDI**

Click on the analyte to view Graphical Data

**All Analyser Result** **Print**

**Date of Result Entered : 17/03/2023**

**Date of Report Published : 04/04/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		No. of Values	SDI	
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	86 mg/dL	-0.80	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	117 mg/dL	1.05	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**Links**

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

  
**MEDICAL OFFICER**  
 PRIMERY HEALTH CENTER  
 PERIYA KAYAPAKKAM

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Periakayapakkam  
 Lab No 12242  
 Month April  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/04/2023


Date of Report Published : 06/05/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI
						CV	St	St		
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	112 mg/dL	-1.03	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.6 mg/dL	2.51	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	118 mg/dL	1.10	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Logout

  
**MEDICAL OFFICER**  
 PRIMARY HEALTH CENTER  
 PERIYA KAYAPAKKAM

# VIEW LAB MONTHLY SUMMARY

**Lab Name** Govt. PHC, Periakayappakkam  
**Lab No** 12242  
**Month** May  
**Year** 2023  
**Constituent Group** Chemistry III

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

Click on the analyte to view Graphical Data

Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	179 mg/dL	-0.20	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">New Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

*S. Sankar*  
 7/6/23  
**MEDICAL OFFICER**  
 PRIMERY HEALTH CENTER  
 PERIYA KAYAPAKKAM



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

<b>Lab Name</b>	<b>Govt. PHC, Periakayapakkam</b>	<b>Lab No</b>	<b>12242</b>
<b>Constituent Group</b>	<b>Chemistry III</b>	<b>Date of Result Entered :</b>	<b>17/06/2023</b>
<b>PT item</b>	<b>Lyophilized human serum based</b>	<b>Date of Report Published :</b>	<b>04/07/2023</b>

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	297 mg/dL	0.70	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	118 mg/dL	0.84	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. PHC, Periakayapakkam  
THE MEDICAL OFFICER, CHITHAMUR  
Chengalpattu  
TAMILNADU

**Coordinator Contact Details:**  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*

*S. Senthil*  
*10.7.23*  
**MEDICAL OFFICER**  
PRIMARY HEALTH CENTER  
PERIYA KAYAPAKKAM

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Periakayapakkam

Details About Robust Analysis

Lab No 12242

Detail About Monthly Summary

Month July

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	106 mg/dL	-0.55	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.6 mg/dL	-0.66	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	119 mg/dL	1.29	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

### Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

*Deepa*  
**MEDICAL OFFICER**  
 PRIMERY HEALTH CENTER  
 PERIYA KAYAPAKKAM

**VIEW LAB MONTHLY SUMMARY**

**Lab Name** Govt. PHC, Periakayapakkam  
**Lab No** 12242  
**Month** August  
**Year** 2023  
**Constituent Group** Chemistry III

**Details About Robust Analysis**

**Detail About Monthly Summary**

**Detail about SDI**

Click on the analyte to view Graphical Data

**All Analyser Result**

**Print**

**Date of Result Entered : 16/08/2023**

**Date of Report Published : 01/09/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	159 mg/dL	-1.84	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	111 mg/dL	-0.60	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**Links**

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

**MEDICAL OFFICER**  
 PRIMERY HEALTH CENTER  
 PERIYA KAYAPAKKAM