

No	Product	Standard Set / Calibrated Value	Expected Value	Observed Value	Problem Observed
	Digital BP Apparatus (BPL B16)	SYS 120 DIA 80 Simulated	± 2%	SYS 115 DIA 77	NIL
13	Semi Auto Analyser ✓	54% RH	>20M	101M	NIL
14	Centrifuge ✓		3000 RPM	2999.6	NIL
15	Refrigerator ✓	25°C	2 °C to 8°C	4°C	
16	Microscope ✓		1000	999.96	NIL
17	Pipette 10-50 ml ✓ 10-100 ml ✓ 100-1000 ml ✓	ISO 8655-6:2002	10 30 50	9.94 29.97 49.94	NIL

All the above equipment are in Working Condition YES  NO


Has the Calibration Passed for all above equipment YES  NO

Calibration Performed Date: 22/03/2023

Next Calibration Due Date: 21/03/2024

For REPUBLIC HEALTHCARE

Certified Engineer name: Mugilarasan A

  
Managing Director

Date / Place: .....

Signature: .....