



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name: Govt. UPHC, Mariyanathapuram
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based
 Lab No: 10693
 Date of Result Entered: 20/07/2023
 Date of Report Published: 07/08/2023

S/No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic	1876	109.35	5.55	6.07	110 mg/dL	0.11	0.28	
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	47	52.80	7.28	3.85	54 mg/dL	0.31	1.12	
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic	1850	5.83	6.00	0.35	5.1 mg/dL	-2.09	0.02	
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic	1776	113.00	4.12	4.66	121 mg/dL	1.72	0.22	
SDI Range				Interpretation							
Within -1.00 to +1.00				Excellent							
Within ±1.01 to ±2.00				Good.							
Within ±2.01 to ±2.99				Accept with caution, Warning Signal.							
Beyond ±3.0				Unacceptable performance. Action Signal.							

தலைமை மருத்துவ அலுவலர்
 டாக்டர் ஆரம்ப சுகாதார நிலையம்,
 மரியாதாபுரம்,
 திருச்செங்கல் மாவட்டம், தஞ்சை.

LAB ADDRESS :
 Govt. UPHC, Mariyanathapuram
 THE MEDICAL OFFICER, MARIYANATHAPURAM
 Mariyanathapuram
 TAMILNADU624003

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Coordinator Contact Details:
 Email: cllnqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****