

VIEW LAB MONTHLY SUMMARY

Lab Name: GOVT PRIMARY HEALTH CENTRE, M.M.KOVILUR
 Lab No: 12141
 Month: April
 Year: 2022
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/04/2022

Date of Report Published : 05/05/2022

SL.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1781	263.81	7.17	18.93	226 mg/dL	-2.00	0.90
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1758	4.67	9.43	0.44	4.8 mg/dL	0.29	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1741	115.66	6.19	7.16	118 mg/dL	0.33	0.34
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

MEDICAL OFFICE
 GOVT. Primary Health Centre
 M.M.KOVILUR
 PINDICULLI
S. Srinivasulu
09/03/22



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name Govt. PHC, M.M.Kovilur

Lab No

12141

Constituent Group Chemistry III

Date of Result Entered :

19/05/2023

PT item Lyophilized human serum based

Date of Report Published :

05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	206 mg/dL	2.67	0.44
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	163	27.91	13.39	3.74	35.9 mg/dL	2.14	0.59
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.5 mg/dL	-1.97	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.27
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ± 1.01 to ± 2.00				Good.						
Within ± 2.01 to ± 2.99				Accept with caution. Warning Signal.						
Beyond ± 3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, M.M.Kovilur
THE MEDICAL OFFICER, DINDIGUL
M.M.Kovilur
TAMILNADU624005

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

MEDICAL OFFICER
GOVT. Primary Health Centre,
M.M.Kovilur,
DINDIGUL.



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name: Govt. PHC, M.M.Kovilur
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No: 12141
 Date of Result Entered: 20/03/2023
 Date of Report Published: 04/04/2023

Sl. No.	Test Name	Method	Analyzer Name	Participants	DV	CV	Participants	Mean Value	SD	SDI	CV
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	90 mg/dL	-0.22	0.31	0.31
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	175	55.20	11.70	6.46	56 mg/dL	0.12	0.98	0.98
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.6 mg/dL	2.23	0.01	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	110 mg/dL	-0.16	0.28	0.28
SDI Range				Interpretation							
Within -1.00 to +1.00				Excellent.							
Within ±1.01 to ±2.00				Good.							
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.							
Beyond ±3.0				Unacceptable performance. Action Signal.							

LAB ADDRESS :
 Govt. PHC, M.M.Kovilur
 THE MEDICAL OFFICER, DINDIGUL
 M.M.Kovilur
 TAMILNADU624005

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

MEDICAL OFFICER
 GOVT. Primary Health Centre,
 M.M.Kovilur,
 DINDIGUL.



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name: Govt. PHC, M.M.Kovilur

Constituent Group: Chemistry III

PT Item: Lyophilized human serum based

Lab No: 12141

Date of Result Entered: 19/08/2023

Date of Report Published: 01/09/2023

Sl. No.	Analysis	Method / Principle / Reagent	Manufacturer/Make	No. of Participants	DV	CV	CV Participants	SD	Your Value	SDI	CV
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	148 mg/dL	-3.06	0.42	0.65
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	174	28.29	15.20	4.30	35 mg/dL	1.56	0.65	0.01
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	115 mg/dL	0.26	0.22	0.22
Interpretation											
SDI Range											
Within -1.00 to +1.00				Excellent.							
Within ±1.01 to ±2.00				Good.							
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.							
Beyond ±3.0				Unacceptable performance. Action Signal.							

LAB ADDRESS :
 Govt. PHC, M.M.Kovilur
 THE MEDICAL OFFICER, DINDIGUL
 M.M.Kovilur
 TAMILNADU624005

Coordinator Contact Details:
 Email: cilingqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Pamela Christudoss

MEDICAL OFFICER
 GOVT. Primary Health Centre,
 M.M.Kovilur,
 DINDIGUL.

M.M.Kovilur



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name Govt. PHC, M.M.Kovilur

Lab No 12141

Constituent Group Chemistry III

Date of Result Entered : 20/06/2023

PT item Lyophilized human serum based

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	283 mg/dL	-0.24	0.69
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	192	32.69	12.96	4.24	33 mg/dL	0.07	0.61
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.52 mg/dL	-0.29	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	111 mg/dL	-0.54	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ± 1.01 to ± 2.00				Good.						
Within ± 2.01 to ± 2.99				Accept with caution. Warning Signal.						
Beyond ± 3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, M.M.Kovilur
THE MEDICAL OFFICER, DINDIGUL
M.M.Kovilur
TAMILNADU624005

Coordinator Contact Details:
Email: cfmqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

MEDICAL OFFICER
GOVT. Primary Health Centre,
M.M.Kovilur,
DINDIGUL.



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name: Govt. PHC, M.M.Kovilur
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No: 12141
 Date of Result Entered: 19/07/2023
 Date of Report Published: 07/08/2023

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	Participants CV	SD	Your Value	SD	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	127 mg/dL	2.91	0.28
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	184	51.41	11.14	5.73	46 mg/dL	-0.94	0.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6 mg/dL	0.49	0.02
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	105 mg/dL	-1.72	0.22
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. PHC, M.M.Kovilur
 THE MEDICAL OFFICER, DINDIGUL
 M.M.Kovilur
 TAMILNADU624005

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

[Handwritten Signature]
 THE MEDICAL OFFICER,
 GOVT. PHC, M.M.KOVILUR
 DINDIGUL