



CALIBRATION CERTIFICATE

Page No: 1 of 1

Certificate No: SBS/CL/23/11594

Customer Name & Address

GOVERNMENT UPGRADED PRIMARY HEALTH CENTRE,
THADDIKKARANKONAM-629851.

SRF No.	SRF/23/00387-0001
SRF Date	20-09-2023
Date of Receipt	20-09-2023
Date of Calibration	20-09-2023
Due Date for Calibration	19-09-2024
Issue Date	21-09-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	100-1000 μ l	Model	FINNPIPETTE F3
Resolution	1 μ l	Material	PVC
Serial Number	PW01675	Operating Range	100-1000 μ l
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature ($^{\circ}$ C)	23.5	Humidity (%RH)	52	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature ($^{\circ}$ C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in μ l

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (\pm)
1	100	99.94	99.96	99.94	99.92	99.93	99.95	-0.05	0.02	0.47
		99.94	99.95	99.96	99.97	99.98				
2	500	499.85	499.86	499.87	499.89	499.86	499.87	-0.13	0.02	0.47
		499.86	499.87	499.89	499.85	499.89				
3	1000	999.96	999.97	999.98	999.96	999.95	999.96	-0.04	0.01	0.47
		999.98	999.94	999.96	999.95	999.95				

Remarks

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

C. Sivabalan
(Calibration Engineer)
C.SIVABALAN



Authorised by:

C. Sivabalan
(Quality Manager/Chief Executive)
C.SIVABALAN