

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Pudhuvalasai  
 Lab No 9261  
 Month June  
 Year 2023  
 Constituent Chemistry III  
 Group

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

[Click on the analyte to view Graphical Data](#)

All Analyser Result

Print

Date of Result Entered : 20/06/2023


Date of Report Published : 04/07/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	302.6 mg/dL	1.08	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	114.7 mg/dL	0.19	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

  
 MEDICAL OFFICER,  
 Govt. Primary Health Centre  
 PUDUVALASAI,  
 Ramnad Dist. -623 522.