

**CALIBRATION / TESTING REPORT**

CUSTOMER DETAILS:		CERTIFICATE NO		TICS/202
The Community Health Centre, Kannuriputhur		EQUIPMENT NAME		MICROPI
Vernacular-626204		MANUFACTURE		THE ICM
CALIBRATION DATE		SERIAL NO		NA
23/05/2023		INSTALL'D		PWB355
CALIBRATION DUE DATE		DEPARTMENT		LAB
22/05/2024				

- 1. Have more dust in machine outside  Yes  No
- 2. Any Physical damaged in machine  Yes  No
- 3. Machine inside Cleaning (if required)  Yes  No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSP122/07/884-01

**TEST REPORT**

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE
VOLUME	µl	10	10.1	+1
		99	99.3	+1

\*The above mentioned instrument has been tested using standards manufactures recommended protocols, having traceability to National/International standards.

ACCEPTABLE FOR USE  CORRECTIVE MAINTENANCE REQUIRED  REMOVED  
 Authorized

Calibrated By: *[Signature]*

MICROPIPETTE (10-100 µl)

Manufacturer :- Thermo scientific  
 Model :- RA  
 Serial no :- PW03357

Micro pipette Calibration on  
 c no :- TICS/2023/HSP/Mrx/01  
 Calibration on :- 23/05/2023  
 Calibration du :- 22/05/2023

4

திருச்சிபி சுகாதாரம்  
 கிழக்கு மருத்துவ சிகிச்சை  
 மருத்துவமனை  
 வி.எஸ். 370