



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name: Govt. PHC, Pudhuvalasai      Lab No: 9261  
 Constituent Group: Chemistry III      Date of Result Entered: 20/04/2023  
 PT item: Lyophilized human serum based      Date of Report Published: 06/05/2023

Sl. No.	Test Name	Instrument Name	Instrument Model	No. of Participants	CV	Participant		CV	SD	CV	SD
						CV	SD				
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	121.6 mg/dL	0.30	0.34	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.09 mg/dL	-0.40	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	114.9 mg/dL	0.57	0.28	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. PHC, Pudhuvalasai

THE MEDICAL OFFICER, GOVT.PRIMARY HEALTH CENTRE, PUDHUVALASAI, RAMANATHAPURAM TALUK, RAMANATHAPURAM

Pudhuvalasai

TAMILNADU623522

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
**MEDICAL OFFICER,**  
 Govt. Primary Health Centre  
 PUDHUVALASAI,  
 Ramnad Dist. - 621 522.