

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt Block PHC, Gudimangalam	Details About Robust Analysis
Lab No	7814	Detail About Monthly Summary
Month	JUNE	Detail about SDI
Year	2023	
Chemistry	Chemistry III	

Click on the analyte to view Graphical Data


Date of Result Entered : 18/06/2023


Date of Report Published : 04/07/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	316 mg/dL	1.98	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	131 mg/dL	3.41	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within =1.01 to ±2.00	Good.
Within =2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond =3.0	Unacceptable performance. Action Signal.

Home	Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Monthly Summary	Change Password	Home	Sign-out	

Site Maintained by
 **Swiftware Solutions**


 Dr. [Name]
 D10/DD18
 -10-