



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Suthamalli
 Lab No 12020
 Month March
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

SL.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Year Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	80.8 mg/dL	-0.83	0.31
2	CREATININE III	Jeffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	6.12	1.33 mg/dL	0.00	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	104.7 mg/dL	-1.07	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



Best Viewed at pixels 1024 X 768

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Suthamsalli
Lab No 12020
Month April
Year 2023
Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participant		Vier Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik / Semi automatic / Fully Automatic	1832	119.45	0.04	7.21	133 mg/dL	1.89	0.34
2	CREATININE III	Jeffes Kinetic - Alkaline Picrate III	Robonik / Semi automatic / Fully Automatic	1820	2.16	8.09	0.18	2.42 mg/dL	1.49	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik / Semi automatic / Fully Automatic	1750	111.61	5.19	5.79	119.7 mg/dL	1.40	0.28

SDI Range	Interpretation
Within -1.00 to -1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Suthamali
 Lab No 12020
 Month May
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered 20/05/2023

Date of Report Published 05/06/2023

Sl. No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Var Value	SDI	S
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robenik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	197.8 mg/dL	1.80	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robenik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.92 mg/dL	1.21	0.01
3	CHOLESTROL III	CHOD-PAP III	Robenik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	110.4 mg/dL	-0.72	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Link			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

(Handwritten Signature)
 மருத்துவ அலுவலர்
 அரசு சுகாதார நிலையம்
 சுகாதாரம் - 621 804.
 மதுரை மாவட்டம்



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Suthamalli
 Lab No 12020
 Month June
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Year Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	231.3 mg/dL	0.32	0.09
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.92 mg/dL	2.05	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.40	5.07	105.5 mg/dL	-1.82	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Suthamalli
 Lab No 12020
 Month July
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023


SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Year Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	3.35	6.07	113.0 mg/dL	0.70	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.33	6.23 mg/dL	1.14	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	120.3 mg/dL	1.37	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Site Maintained by
Swiftware Solutions

Best Viewed at pixels 1024 X 768


 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 சுத்தமல்லி - 621 874.
 அரியலூர் மாவட்டம்

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt PHO, Gudimallai	Details About Robocast Analysis
Lab No	12020	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analysis to view Graphical Data](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

SL.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Vier Woban mg/dL	SDI	UP
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik Semi automatic / Fully Automatic	1841	175.58	5.13	9.00	183.7 mg/dL	0.88	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik Semi automatic / Fully Automatic	1852	1.76	6.77	0.12	1.92 mg/dL	1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik Semi automatic / Fully Automatic	1704	113.80	4.07	4.64	1.14 mg/dL	0.04	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Modify Address	Change Password	Home	Sign-out
Yearly Summary			

R. Jeyapalan

மருத்துவ அலுவலர்
அரசு ஆராய்ச்சி கார்ப்பரேஷன்
கத்தமல்லி - 621 011
அரியலூர் மாளிகை

