

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Palada , M
 Lab No 12478
 Month August
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 28/08/2023

Date of Report Published : 01/09/2023

Sl. No.	Analyte	Method / Principle Name	Analyser Name	No. of Participants	UV	Temperature	Refractive Index	Conductivity	SDI	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	9.13	9.06	178.5 mg/dL	9.34	9.43
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.78	9.77	9.12	9.73 mg/dL	-0.25	9.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.90	4.97	4.94	114.3 mg/dL	9.15	9.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

09/11