

9/21/23, 5:17 PM

External Quality Assurance Scheme - Print Monthly Summary



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name	Govt. PHC, Kalathur .V	Lab No	12498
Constituent Group	Chemistry III	Date of Result Entered :	19/08/2023
PT Item	Lyophilized human serum based	Date of Report Published :	01/09/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	180 mg/dL	0.49	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	105 mg/dL	-1.90	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Kalathur .V
CLINICAL LABORATORY
Kalathur .V
TAMILNADU621117

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. PHC, Kalathur .V
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 12498

Date of Result Entered : 15/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	118 mg/dL	1.43	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.3 mg/dL	-1.51	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	115 mg/dL	0.43	0.22
SDI Range										
				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. PHC, Kalathur .V
 CLINICAL LABORATORY
 Kalathur .V
 TAMILNADU621117

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore.

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components

6/20/23, 4:25 PM

U	0.76	0.06	0.38
---	------	------	------

7/22/23, 11:06 AM

External Quality Assurance Scheme - Print Monthly Summary

External Quality Assurance Scheme - Print Monthly Summary



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name: Govt. PHC, Kalathur .V
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 12498
Date of Result Entered: 20/06/2023
Date of Report Published: 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	287 mg/dL	0.02	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.1 mg/dL	-2.75	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	110 mg/dL	-0.74	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Kalathur .V
CLINICAL LABORATORY
Kalathur .V
TAMILNADU621117

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

will Rectify next time

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

shubh

6/20/23, 4:25 PM

0.76	0.06	0.38
------	------	------

External Quality Assurance Scheme - Print Monthly Summary



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



Lab Name: Govt. PHC, Kalathur .V
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No: 12498
 Date of Result Entered: 20/05/2023
 Date of Report Published: 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	180 mg/dL	-0.09	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	118 mg/dL	0.58	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Kalathur .V
 CLINICAL LABORATORY
 Kalathur .V
 TAMILNADU621117

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

Handwritten Signature
 மருத்துவ சேவைகள்
 அரசு ஆரம்ப சுகாதார நிலையம்,
 வ. களத்தூர் - 621 117.
 பெரியவாழை மாவட்டம்.



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

U	0.76	0.06	0.38
---	------	------	------

Lab Name Govt. PHC, Kalathur .V

Lab No 12498

Constituent Group Chemistry III

Date of Result Entered : 17/04/2023

PT item Lyophilized human serum based

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	140 mg/dL	2.85	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	111 mg/dL	-0.11	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Kalathur .V
CLINICAL LABORATORY

will Rectify next time

மருத்துவ அலுவலர்
ஆரச் ஆரம்ப சுகாதாய நிலையம்,
வ. களத்தூர் - 824 117.
முகாம்மது ஸாஹீப்



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. PHC, Kalathur .V

Constituent Group Chemistry III

PT Item Lyophilized human serum based

Lab No

12498

Date of Result Entered :

11/03/2023

Date of Report Published :

04/04/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	77 mg/dL	-2.12	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	103 mg/dL	-1.37	0.28
SDI Range										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Kalathur .V
CLINICAL LABORATORY
Kalathur .V
TAMILNADU621117

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

LD (E) 10/1/2023
அரசு ஆரம்ப சுகாதார நிலையம்,
வி. கலாத்தூர் - 621 117.
மொழிபெயர்ச்சி இலாபம்.

home.cmcvellore.ac.in/clinqc/NewMonthlySummary/printResult_new.aspx?resDetId=263&noSamples=1&month=3&year=2023&groupid=9&yearGroup=110
Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential