



### CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 17.04.2023
		Calibration valid: 16.10.2023
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/s, PRIMARY HEALTH CENTRE, SAMATHUR, COIMBATORE.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMOSCEINTIFIC
ii	Model	FINNPIPETTE F3
iii	Value	10 µl -100 µl
iv	Serial No	RW02672

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2023	06-OCT-2024

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems,

(Authorized signatory) 