



| CALIBRATION REPORT | |
|---|--|
| Report No : HBMS/CAL/23-24 | Calibration Date : 17.04.2023 Calibration valid: 16.10.2023 |
| 1.1.CUSTOMER DETAILS: | |
| Name and Address of the organization | M/s, PRIMARY HEALTH CENTRE, SAMATHUR, COIMBATORE. |
| 1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT): | |
| EQUIPMENT NAME | MICRO PIPETTE |
| i Manufactured by | THERMOSCIENTIFIC |
| ii Model | FINNPIPETTE |
| iii Value | 100µl-1000µl |
| iv Serial No | RW02672 |

Calibration standards used :

| Master instrument | Model | Serial no | Date of calibration | Calibration due on |
|--------------------------|---------|------------|---------------------|--------------------|
| Electrical compact scale | SF-400C | SF400-2240 | 07-OCT-2023 | 06-OCT-2024 |

Calibration test Report.

| Number of measurement | Channel 1 (Volume) | | |
|-----------------------|--------------------|--------|---------|
| | 100µl | 500µl | 1000µl |
| 1 | 100.02 | 500.05 | 1000.15 |
| 2 | 100.03 | 500.09 | 1000.13 |
| 3 | 100.06 | 500.03 | 1000.12 |
| 4 | 100.08 | 500.08 | 1000.08 |
| 5 | 100.00 | 499.95 | 1000.09 |
| 6 | 100.09 | 500.09 | 1000.07 |

For Heart Beat Medical Systems,

(Signature)
(Authorized signatory) 