



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name	Govt. UPHC, Kodikalpalayam	Lab No	10911
Constituent Group	Chemistry III	Date of Result Entered :	20/06/2023
PT item	Lyophilized human serum based	Date of Report Published :	04/07/2023

Sl. No.	Analyte	Method / Principle Name	Analyzer Model	No. of Participants	DV	Participants		Your Value	SDI	
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	298 mg/dL	0.77	0.69
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	192	32.69	12.96	4.24	32.1 mg/dL	-0.14	0.61
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	111.8 mg/dL	-0.38	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UPHC, Kodikalpalayam
THE MEDICAL OFFICER, SOUTH STREET, KODIKALPALAYAM
Thiruvarur Kodikalpalayam
TAMILNADU610001

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

S. G. Vasudhayan
MEDICAL OFFICER,
GOVT. URBAN PRIMARY,
HEALTH CENTRE,
TIRUVARUR.