



**Vanavil Calibrations**

Make It Simple, But Significant !

## CALIBRATION CERTIFICATE

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> GOVERNMENT PRIMARY HEALTH CENTRE RAMACHANDRAPURAM - 642 120	Report No	VCPL/2023/003-5
	Date of Issue :	24/Mar/2023
	Date of Receipt :	23/Mar/2023
	Date of Calibration :	23/Mar/2023
	Cal Due Date :	22/Mar/2024
	Calibration at	LAB

### Details of Device Under Calibration:

Description:	MICRO PIPETTE	Make :	PRIMAN
Range:	100µl-1000µl	Model/Type:	
Resolution:	10µl	Accuracy :	As per Manual
Serial Number:	111652067	Location:	LABORATORY
ID number:		Condition of DUC:	GOOD

Temperature	22.8 °C	National/International standard:	-
Humidity	52 % Rh	Cal procedure no:	-

### Master Details

SL no:	Description	Serial No	Certificate no	Valid Till
1	ELECTRONIC BALANCE	16031172	E150323-1	15-Mar-24


### CALIBRATION RESULT

S.NO	DUC Readings (µl )	Standard Readings Mean Value ( mg )	ERROR (mg)
1	100	99.84	0.16
2	500	499.91	0.09
3	1000	999.87	0.13

### Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

  
T.SUDHAHAR  
(CALIBRATION ENGINEER)

Authorised by:

  
**S. LAKSHMANAN MURUGAN,**  
Technical Manager  
**Vanavil Calibrations Pvt Ltd.**

