



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name: Govt. Block PHC, Perumanallur Lab No: 4593
 Constituent Group: Chemistry III Date of Result Entered: 19/08/2023
 PT item: Lyophilized human serum based Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	177 mg/dL	0.16	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.97 mg/dL	1.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	114 mg/dL	0.04	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Perumanallur
 GOVT. PRIMARY HEALTH CENTRE TIRUPUR _ D.T.,
 Perumanallur
 TAMILNADU641666

Coordinator Contact Details:
 Email: clinqc@cmvellore.ac.in
 Contact Number: 0416-2283102

M. N. Srinivasan
BLOCK MEDICAL OFFICE
 Govt. Primary Health Centre
 Perumanallur, Tirupur - 641 666.

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name Govt. Block PHC, Perumanallur

Lab No

4593

Constituent Group Chemistry III

Date of Result Entered :

20/07/2023

PT item Lyophilized human serum based

Date of Report Published :

07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	108 mg/dL	-0.22	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.47 mg/dL	1.83	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	122 mg/dL	1.93	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Perumanallur
GOVT. PRIMARY HEALTH CENTRE TIRUPUR _D.T.,
Perumanallur
TAMILNADU641666

M. Per
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE,
PERUMANALLUR - 641 666.
TIRUPUR DISTRICT

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name: Govt. Block PHC, Perumanallur
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

4593

19/06/2023

04/07/2023

Analyser: Robonik
 Participants: 1823, 1848, 1804

Item No	Test Name	Method	Participant No	Result	Reference Range	Unit
1	GLUCOSE III	GOD-POD III	1823	286.63	14.82	296 mg/dL
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1848	2.57	0.17	2.8 mg/dL
3	CHOLESTROL III	CHOD-PAP III	1804	113.73	5.07	114 mg/dL

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. Block PHC, Perumanallur
 GOVT. PRIMARY HEALTH CENTRE TIRUPUR _ D.T ,
 Perumanallur
 TAMILNADU641666

MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTER
PERUMANALLUR - 641 666.
TIRUPUR DISTRICT

Dr. Pamela Christudoss
 17/7/2023
 Dist. Medical Officer
 Perumalapur

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components