

## CALIBRATION CERTIFICATE

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> GOVERNMENT PRIMARY HEALTH CENTRE RAMACHANDRAPURAM - 642 120	Report No	VCPL/2023/003-4
	Date of Issue :	24/Mar/2023
	Date of Receipt :	23/Mar/2023
	Date of Calibration :	23/Mar/2023
	Cal Due Date :	22/Mar/2024
	Customer Reference	Telephone Communication
	Calibration at	LAB

**Details of Device Under Calibration:**

Description:	MICRO PIPETTE	Make :	MICROLUX
Range:	5µl-50µl	Model/Type:	
Resolution:	-	Accuracy :	As per Manual
Serial Number:	GPHC/PIP/02	Instrument Location:	LABORATORY
ID number:		Condition of DUC:	GOOD

Temperature	22.8 °C	National/International standard:	-
Humidity	52 % Rh	Cal procedure no:	-

**Master Details**

SL no:	Description	Serial No	Certificate no	Valid Till
1	ELECTRONIC BALANCE	16031172	E150323-1	15-Mar-24


**CALIBRATION RESULT**

S.NO	DUC Readings (µl )	Standard Readings Mean Value ( mg )	ERROR (mg)
1	5	4.92	0.08
2	20	19.86	0.14
3	50	49.88	0.12

**Remarks:**

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

  
**T.SUDHAHAR**  
 (CALIBRATION ENGINEER)

Authorised by:

  
**S. LAKSHMANAN MURUGAN,**  
 Technical Manager  
**Vanavil Calibrations Pvt Ltd.**

