

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: GOVERNMENT PRIMARY HEALTH CENTRE RAMACHANDRAPURAM - 642 120	Report No	VCPL/2023/003-3
	Date of Issue :	24/Mar/2023
	Date of Receipt :	23/Mar/2023
	Date of Calibration	23/Mar/2023
	Cal Due Date :	22/Mar/2024
	Calibration at	LAB

Details of Device Under Calibration:

Description:	MICRO PIPETTE	Make :	-
Range:	FIXED 10 μ l	Model/Type:	RIGHTQPLUS
Resolution:	-	Accuracy :	As per Manual
Serial Number:	GPHC/PIP/01	Instrument Locatio	LABORATORY
ID number:		Condition of DUC:	GOOD

Temperature	22.8 °C	National/International standard:	-
Humidity	52 % Rh	Cal procedure no:	-

Master Details

SL no:	Description	Serial No	Certificate no	Valid Till
1	ELECTRONIC BALANCE	16031172	E150323-1	15-Mar-24


CALIBRATION RESULT

S.NO	DUC Readings (μ l)	Standard Readings Mean Value (mg)	ERROR (mg)
1	10	9.94	0.06


Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:


T.SUDHAKAR
 (CALIBRATION ENGINEER)

Authorised by:




S.LAKSHMANAN MURUGAN,
 Technical Manager
Vanavil Calibrations Pvt Ltd.