


CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 17.04.2023
	Calibration Valid: 16.10.2023
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/s, PRIMARY HEALTH CENTRE, SAMATHUR, COIMBATORE.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	PIPETTE
I   Manufactured by	DIATEK
II   Model	10µl
III   Serial No	190290097

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	10µl
1	10.05
2	10.06
3	10.01
4	9.97
5	10.01
6	10.02

For Heart Beat Medical Systems,



(Authorized signatory)

