


CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 17.04.2023
	Calibration Valid: 16.10.2023
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/s, PRIMARY HEALTH CENTRE, SAMATHUR, COIMBATORE.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	PIPETTE
I Manufactured by	MICROLUX
II Model	5µl - 50µl
III Serial No	SM/PHC/LAB/PIP/001

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	05µl	20µl	50µl
1	05.05	20.08	50.97
2	05.06	20.07	20.02
3	05.01	20.05	20.96
4	05.97	20.99	20.02
5	05.01	20.08	20.09
6	05.02	20.95	20.04

For Heart Beat Medical Systems



(Authorized signatory)

