

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: GOVT PRIMARY HEALTH CENTRE, POOLAVADI - 642 206.	Report No :	VCPL/2023/004-01
	Date of Issue :	23/Mar/2023
	Date of Receipt :	24/Mar/2023
	Date of Calibration :	23/Mar/2023
	Cal Due Date :	22/Mar/2024
	Calibration at	IN LAB

Details of Device Under Calibration:

Description:	MICRO PIPETTE	Make :	-
Range:	10-100µl	Model/Type:	HANDYPETTE
Resolution:	1µl	Accuracy :	-
Serial Number:	20018362	Instrument Location:	LAB
ID number:	-	Condition of DUC:	GOOD
Environmental Conditions		Calibration Method Used	
Temperature	22.8 °C	National/International standard:	-
Humidity	52 % Rh	Cal procedure no:	-

Master Details

SL no:	Description	Serial No	Certificate no	Valid Till
1	ELECTRONIC BALANCE	16031172	E150323-1	15/Mar/24

Result of calibration

Sl no:	DUC Readings (µl)	Standard Readings Mean Value (mg)	ERROR (mg)
1	10	9.93	0.07
	50	49.88	0.12
3	100	99.79	0.21

1. DUC is Defined as Device Under Calibration.
2. Calibration Frequency decided by the customer.
3. The above instrument is working in good condition.

Calibrated By:



M. BALAJI


(CALIBRATION ENGINEER)

Authorised by:



M. KANNAN,
Technical Manager

Vanavil Calibrations Pvt Ltd.



Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indira Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

Web : vanavilcalibrations.com | Email : vanavilcalibrations@gmail.com