

17/07/23 12:32:09

TEST NAME: CHOLE

Using Stored Value

Blk Res: 0.040

Factor: 500.000

Sr.No: 1

ABS: 0.223

Res: 111.4593mg/dl LO

Sr.No: 2

Init.OD: 0.543 HI

Delta: 0.250

Res: 5.7748 HI

Sr.No: 3

ABS: 0.461

Res: 106.9965mg/dl NM

Sr.No: 4

ABS: 0.309

Res: 71.7274mg/dl NM

# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Govt. PHC, Ganapathipalayam

Biochemistry III

Lyophilized human serum based

Lab No

12207

Date of Result Entered :

17/07/2023

Date of Report Published :

07/08/2023

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
				CV	SD			
GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	106.99 mg/dL	-0.39	0.28
2 CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1850	5.83	6.00	0.35	5.77 mg/dL	-0.17	0.02
3 CHOLESTROL III	CHOD-PAP III	1776	113.00	4.12	4.66	111.45 mg/dL	-0.33	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Ganapathipalayam

THE MEDICAL OFFICER, MODAKURICHI

Ganapathipalayam

TAMILNADU638152

Coordinator Contact Details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

மருத்துவ அலுவலர்  
அரசு ஆரம்ப சுகாதார நிலை  
கணபதிபாளையம்

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

April-23

11:06:12

20/04/23

TEST NAME: GLUCOSE  
Using Stored Value  
Blk Res: 0.050  
Using Stored Value

Std Res: 0.433  
Factor: 231.200

St. No: 1

ABS: 0.437

Res: 101.0999mg/dl NM

I No: 2

BS: 0.535

Res: 123.6171mg/dl HI

Factor: 454.564

I No: 1

S: 0.234

S: 106.4671mg/dl LO

No: 2

0.237

107.6107mg/dl LO

**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Govt. PHC, Ganapathipalayam  
Chemistry III  
Lyophilized human serum based

Lab No: 12207  
Date of Result Entered: 20/04/2023  
Date of Report Published: 06/05/2023

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
				CV	SD				
GOD-POD III	Robonik ( Semi automatic )	1832	119.45	6.04	7.21	123.61 mg/dL	0.58	0.34	
Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	1.84 mg/dL	-1.83	0.01	
CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	107.61 mg/dL	-0.69	0.28	
Interpretation									
Within $\pm 1.01$ to $\pm 2.00$	Excellent.								
Within $\pm 2.01$ to $\pm 2.99$	Good.								
Beyond $\pm 3.0$	Accept with caution. Warning Signal. Unacceptable performance. Action Signal.								

LAB ADDRESS :  
Govt. PHC, Ganapathipalayam  
THE MEDICAL OFFICER, MODAKURICHI  
Ganapathipalayam  
TAMILNADU638152

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*Pamela Christudoss*

மருத்துவ அலுவலர்  
அரசு ஆரம்ப சுகாதார நிலையம்  
கணபதியூர்

*N. Anand*

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential

20/05/23 12:06:02

## CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023

May-23



PC-1024

TEST NAME: GLUCOSE  
Using Stored Value  
Blk Res: 0.020  
Using Stored Value  
Std Res: 0.414  
Factor: 241.654  
Sr.No: 1  
ABS: 0.763  
Res: 184.4130mg/dl HI  
Cr. No: 2  
Init.OD: 0.468 HI  
Delta: 0.079  
Res: 1.7913 HI

Govt. PHC, Ganapathipalayam

Lab No

12207

Chemistry III

Date of Result Entered :

20/05/2023

Unphosphorized human serum based

Date of Report Published :

05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	184.41 mg/dL	0.38	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.79 mg/dL	0.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	106.77 mg/dL	-1.34	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. PHC, Ganapathipalayam  
THE MEDICAL OFFICER, MODAKURICHI  
Ganapathipalayam  
TAMILNADU638152

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*N. Anbum*  
மருத்துவ அலுவலர்  
அரசு ஆரம்ப சுகாதார நிலையம்  
கணபதிபாளையம்

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

20/06/23 12:12:18

External Quality Assurance Scheme - Print Monthly Summary

June-23



PC-1024

TEST NAME CREAT  
Using Stored Value  
Std Res: 0.007  
Factor: 23.108  
Sr No: 1  
Init.OD: 0.413 HI  
Delta: 0.103  
Res: 2.3816 HI  
ABS: 0.304  
Res: 109.2375mg/dl LO  
ABS: 1.107  
Res: 285.1249mg/dl HI

# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023

Govt. PHC, Ganapathipalayam

Lab No

12207

Chemistry III

Date of Result Entered :

20/06/2023

Lyophilized human serum based

Date of Report Published :

04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	285.12 mg/dL	-0.10	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.38 mg/dL	-1.11	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	109.23 mg/dL	-0.89	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. PHC, Ganapathipalayam  
THE MEDICAL OFFICER, MODAKURICHI  
Ganapathipalayam  
TAMILNADU638152

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*N. Anbu*  
மருத்துவ அலுவலர்  
அரசு ஆரம்ப சுகாதார நிலையம்  
கணபதிபாளையம்

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential

17/07/23 12:32:09

TEST NAME: CHOLE

Using Stored Value

Blk Res: 0.040

Factor: 500.000

Sr.No: 1

ABS: 0.223

Res: 111.4593mg/dl LO

Sr.No: 1

Init.OD: 0.543 HI

Delta: 0.250

Res: 5.7748 HI

Sr.No: 3

ABS: 0.461

Res: 106.9965mg/dl NM

Sr.No: 4

ABS: 0.309

Res: 71.7274mg/dl NM

## CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Govt. PHC, Ganapathipalayam

Ministry III

Lithium human serum based

Lab No

12207

Date of Result Entered :

17/07/2023

Date of Report Published :

07/08/2023

	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U	
					CV	SD				
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	106.99 mg/dL	-0.39	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.77 mg/dL	-0.17	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	111.45 mg/dL	-0.33	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. PHC, Ganapathipalayam

THE MEDICAL OFFICER, MODAKURICHI

Ganapathipalayam

TAMILNADU638152

Coordinator Contact Details:

Email:clinq@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

மருத்துவ அலுவலர்  
அரசு ஆரம்ப சுகாதார நிலையம்  
கணபதிபாளையம்

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Ganapathipalayam  
 Lab No 12207  
 Month August  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 18/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	179.42 mg/dL	0.43	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.61 mg/dL	-1.26	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	108.6 mg/dL	-1.12	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

மருத்துவ அலுவலர்  
 அரசு ஆரம்ப சுகாதார நிலையம்  
 கணபதிபாளையம்

Links