

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Velur Lab No 12494

Constituent Group Chemistry III Date of Result Entered : 10/06/2023

PT item Lyophilized human serum based Date of Report Published : 04/07/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	CDI	
						CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	299 mg/dL	0.83	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	118 mg/dL	679	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	2.9 mg/dL	-21.86	0.24

SDI Range	Interpretation		
Within -1.00 to +1.00	Excellent.		
Within ±1.01 to ±2.00	Good.		
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.		
Beyond ±3.0	Unacceptable performance. Action Signal.		

LAB ADDRESS :

Govt. PHC, Velur CLINICAL LABORATORY Velur TAMILNADU

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

********* End of Report *******