



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name Govt. PHC, 4 Veerapandi

Lab No 12054

20/06/2023

Constituent Group Chemistry III

Date of Result Entered :

04/07/2023

PT Item Lyophilized human serum based

Date of Report Published :

04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	284 mg/dL	-0.18	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.9 mg/dL	1.93	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	118 mg/dL	0.84	0.24
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

### LAB ADDRESS :

Govt. PHC, 4 Veerapandi  
THE MEDICAL OFFICER, NO 4 VEERAPANDI  
Coimbatore  
TAMILNADU 641019

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*[Signature]*  
மருத்துவ அலுவலர்  
சென்னை -  
சென்னை, கோவை -

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

