



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - FEBRUARY 2023



PC-1024

Lab Name Govt. Block PHC, Mangalam

Lab No

7522

Constituent Group Chemistry III

Date of Result Entered :

22/02/2023

PT item Lyophilized human serum based

Date of Report Published :

08/03/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1868	155.11	6.56	10.18	146 mg/dL	-0.89	0.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1872	3.40	9.00	0.31	3.9 mg/dL	1.63	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1788	112.79	5.23	5.90	121 mg/dL	1.39	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

## LAB ADDRESS :

Govt. Block PHC, Mangalam  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVANNAMALAI DISTRICT.  
Mangalam  
TAMILNADU606752

Coordinator Contact Details:  
Email:clingc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*APX*

Homogeneity and Stability of the sample is passed.

MEDICAL OFFICER  
GOVT. UPGRADED PRIMARY HEALTH CENTER  
MANGALAM. 606 752, T V MALAI DISTRICT



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. Block PHC, Mangalam

Lab No

7522

Constituent Group Chemistry III

Date of Result Entered :

20/03/2023

PT Item Lyophilized human serum based

Date of Report Published :

04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	93.2 mg/dL	0.25	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.35 mg/dL	0.17	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	115.2 mg/dL	0.74	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within $\pm 1.01$ to $\pm 2.00$				Good.						
Within $\pm 2.01$ to $\pm 2.99$				Accept with caution. Warning Signal.						
Beyond $\pm 3.0$				Unacceptable performance. Action Signal.						

### LAB ADDRESS :

Govt. Block PHC, Mangalam

THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVANNAMALAI DISTRICT, Mangalam

TAMILNADU606752

Coordinator Contact Details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

*[Signature]*

Homogeneity and Stability of the sample is passed.

MEDICAL OFFICER

GOVT. UPGRADED PRIMARY HEALTH CENTRE,  
MANGALAM - 606 752, T V MANAI DISTRICT