



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name: Govt. Block PHC, Kuruvikulam
Lab No: 7595
Constituent Group: Chemistry III
Date of Result Entered: 19/08/2023
PT item: Lyophilized human serum based
Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	172 mg/dL	-0.40	0.42
2	CREATININE III	Jaffes Kinetec - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	112 mg/dL	-0.39	0.22
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
Govt. Block PHC, Kuruvikulam
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, TIRUNELVELU DISTRICT.
Kuruvikulam
TAMILNADU627754

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christodoss
Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

M.F.H.
Medical Officer
Govt. Primary Health Centre
Kuruvikulam - 627754

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components