



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - FEBRUARY 2023



PC-1024

Lab Name: Govt. Block PHC, Mangalam
Constituent Group: Chemistry III
PT item: Lyophilized human serum based

Lab No: 7522
Date of Result Entered: 22/02/2023
Date of Report Published: 08/03/2023

SlNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	155.11	6.56	10.18	146 mg/dL	-0.89	0.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1872	3.40	9.00	0.31	3.9 mg/dL	1.63	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1788	112.79	5.23	5.90	121 mg/dL	1.39	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
Govt. Block PHC, Mangalam
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVANNAMALAI DISTRICT,
Mangalam
TAMILNADU606752

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

PC

Homogeneity and Stability of the sample is passed.
MEDICAL OFFICER
GOVT. UPGRADED PRIMARY HEALTH CENTRE,
MANGALAM. 606 752, T V MALAI DISTRICT