

# S J V Scientific Company

20, Railway Lane Street, 1<sup>st</sup> Cross, Chekkan Thottam, Ganapathy, Coimbatore – 641 006.

E-mail: [sjv.scientific@gmail.com](mailto:sjv.scientific@gmail.com), Ph no: 97866-90789

GSTIN: 33AJXPR6161K1ZG

Ref-2023/Cal/PHC/VP/010/01

To,

The Medical Officer,  
Govt Primary Health Centre,  
No 4, Veerapandi, Naikanur,  
Coimbatore – 641 019.

## SEMI AUTO ANALYZER CALIBRATION CERTIFICATE

### Calibrated Data:

- Lamp output found OK
- Optical Unit found OK
- Filter Voltages (in range 2-10V)
- Peristaltic pump & aspiration Volume OK
- Results found OK

Model : prietest TOUCH  
Serial No : ATCD0100519RBK  
Power : 100 VA  
Voltage : 115-230VAC, 50-60Hz

Comments: This above-mentioned product is working as expected.

Recommendation: -NIL-

Calibration Date : 20.09.2023

Due Date : 19.09.2024

SJV Scientific Company



Sales & Service:

Blood Glucometer - BP Apparatus - Weighing Scale - Nebulizer - ECG Machine - Lab Kits & Chemicals - Surgical Instruments.



Scanned with OKEN Scanner