

S J V Scientific Company

20, Railway Lane Street, 1st Cross, Chekkan Thottam, Ganapathy, Coimbatore – 641 006.

E-mail: sjv.scientific@gmail.com. Ph no: 97866-90789

GSTIN: 33AJXPR6161K1ZG

Ref -2023/Cal/PHC/VP/017/01

To,

The Medical Officer,
Govt Primary Health Centre,
No 4, Veerapandi, Naikanur,
Coimbatore – 641 019.

CALIBRATION REPORT

DR. CENTRIFUGE MACHINE 8 TUBES CALIBRATION:

- Fitted with resiliently mounted motor for vibration free performance.
- A specially wound 5 step transformer is used for speed regulation.
- The highly sophisticated elegant body of the equipment is made with fiber re-enforced plastic which gives protection against corrosion and electric shock.

Technical Data:

Maximum Speed :4000 RPM,
Maximum RCF:1880 x g,
Motor: 1/20 HP AC/DC universal.

Comments: This above mentioned product is working as expected.

Calibration Date : 20.09.2023

Due Date : 19.09.2024

SJV Scientific Company



Sales & Service:

Blood Glucometer - BP Apparatus - Weighing Scale - Nebulizer - ECG Machine - Lab Kits & Chemicals - Surgical Instruments.



Scanned with OKEN Scanner