

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. Block PHC, Peraiyur Lab No 7475

Constituent Date of Result Entered : **Chemistry III** 20/04/2023 Group

PT item Lyophilized human serum based Date of Report Published: 06/05/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	U
31.110						CV	SD	Value	301	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	114.5 mg/dL	-0.69	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.6 mg/dL	2.51	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	118 mg/dL	1.10	0.28

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS:

Govt. Block PHC, Peraiyur

Mr.N.Muniyaraj, Laboratory technician RAMANATHAPURAM DISTRICT.

Coordinator Contact Details:

Peraiyur

TAMILNADU623120

Panela Christudoss Dr. Pamela Christudoss **CMC EQAS Coordinator**

Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Christian Medical College, Vellore

> Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components ****** End of Report *****