



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. UOPHC, Muthupetta

Lab No: 7269

Constituent Group: Chemistry III

Date of Result Entered: 26/08/2023

PT Item: Lyophilized human serum based

Date of Report Published: 01/09/2023

Sl No	Analyte	Method / Principle Name	Analyser (Name)	No of Participants	DV	Participation		Your Result	SD	U
						CV	SO			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.56	192 mg/dL	1.82	5.42
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	174	28.29	15.20	4.30	27 mg/dL	-0.20	0.55
3	CREATININE III	Jaffe's Kinetic - Alkaline Phosphate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	0.77	0.12	1.7 mg/dL	-0.50	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	165 mg/dL	-1.90	0.22
SDI Range Interpretation										
				Within -1.00 to +1.00 Excellent.						
				Within +1.01 to +2.00 Good.						
				Within +2.01 to +2.99 Accept with caution. Warning Signal.						
				Beyond +3.0 Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. UOPHC, Muthupetta
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVARUR DISTRICT
 Muthupetta
 TAMILNADU 614704

Coordinator Contact Details:
 Email: cdhqc@govnetlore.ac.in
 Contact Number: 0416-2893192

Ravida Chidambaram

Dr. Parvathi Christudas
 CMC EQAS Coordinator
 Christian Medical College, Vellore

MEDICAL OFFICE
 Govt Primary Health Center
 MUTHUPET - 614 704
 Thiruvavur District

Lab Name Govt. UGPHC, Muthupettai

Lab No

7269

Constituent Group Chemistry III

Date of Result Entered :

20/07/2023

PT item Lyophilized human serum based

Date of Report Published :

07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Year Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	112 mg/dL	0.44	0.28
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	184	51.41	11.14	5.73	54 mg/dL	0.45	0.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6 mg/dL	0.49	0.02
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	115 mg/dL	0.43	0.22
SDI Range										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. UGPHC, Muthupettai

THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVARUR DISTRICT.

Muthupettai

TAMILNADU614704

Coordinator Contact Details:

Email: clingc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christodoss

Dr. Pamela Christodoss

CMC EQAS Coordinator

Christian Medical College, Vellore

MEDICAL OFFICER
Govt. Primary Health Centre
MUTHUPET - 614 704
Thiruvavarur District



**CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023**



Lab Name: Govt. UGPHC, Muthupettai

Lab No: 7289

Constituent Group: Chemistry III

Date of Result Entered: 13/06/2023

PT Item: Lyophilized human serum based

Date of Report Published: 04/07/2023

JUNE - 2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Year Value	SCI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	282 mg/dL	-0.31	0.89
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.4 mg/dL	-0.99	0.91
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	100 mg/dL	-2.71	0.24
SDI Range										
				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within +1.01 to +2.00				Good.						
Within +2.01 to +2.99				Accept with caution. Warning Signal.						
Beyond +3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

**Govt. UGPHC, Muthupettai
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVARUR DISTRICT.
Muthupettai
TAMILNADU 614704**

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

**Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *******

MEDICAL OFFICER
Govt. Primary Health Centre
MUTHUPETTAI - 614 704
Thiruvavur District



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name: Govt. UGPHC, Muthupettai

Lab No: 7289

Constituent Group: Chemistry III

Date of Result Entered: 20/05/2023

PT item: Lyophilized human serum based

Date of Report Published: 05/06/2023

May-23

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	188 mg/dL	0.76	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	121 mg/dL	1.10	0.27
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
Govt. UGPHC, Muthupettai
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVARUR DISTRICT.
Muthupettai
TAMILNADU 614704

Coordinator Contact Details:
Email: cfmqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore


MEDICAL OFFICE
Govt. Primary Health Centre,
MUTHUPEL - 614 704
Thiruvavur District.

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PG-102A

Lab Name: Govt. UGPHC, Muthupettai

Constituent Group: Chemistry III

PT Item: Lyophilized human serum based

Lab No: 7269

Date of Result Entered: 20/04/2023

Date of Report Published: 06/05/2023

APRIL-23

Sl. No	Analyte	Method / Principle / Reagent	Analyzer Name	No. of participants	CV	Participants CV	Mean value	SD	CV	
1	GLUCOSE III	GOD-POD III	Robonk (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	119 mg/dL	0.06	0.24
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonk (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.2 mg/dL	0.22	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonk (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	117 mg/dL	0.93	0.28
SDI Range										
				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within +1.01 to +2.00				Good.						
Within +2.01 to +2.99				Accept with caution. Warning Signal.						
Beyond +3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS:

Govt. UGPHC, Muthupettai
Muthupettai
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVARUR DISTRICT.
TAMILNADU 614704

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

MEDICAL OFFICER
Govt. Primary Health Centre
MUTHUPET 614 704
Thiruvavur District

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1023

Lab Name: Govt. UGPHC, Muthupettai
 Constituent Group: Chemistry III
 PT Rem: Lyophilized human serum based

Lab No: 7269
 Date of Result Entered: 19/03/2023
 Date of Report Published: 04/04/2023

MARCH - 25

SL/NO	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	86 mg/dL	-0.80	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.2 mg/dL	-1.07	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	114 mg/dL	0.53	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±3.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. UGPHC, Muthupettai
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVARUR DISTRICT.
 Muthupettai
 TAMILNADU614704

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

MEDICAL OFFICER
 Govt. Primary Health Centre
 MUTHUPET - 614 704
 Thiruvannur District