

9/28/2023

External Quality Assurance Scheme - Print Monthly Summary



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name Govt. PHC, Veerampakkam
Constituent Group Chemistry III
PT item Lyophilized human serum based

Lab No 9264
Date of Result Entered : 19/08/2023
Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Biosystems	34	172.03	5.51	9.48	169 mg/dL	-0.32	3.25
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	1.76	6.47	0.11	1.6 mg/dL	-1.40	0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	33	114.98	3.90	4.49	114 mg/dL	-0.22	1.56

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Veerampakkam
THE MEDICAL OFFICER, GOVT. PRIMARY HEALTH CENTER, VEERAMPAKKAM, VANDAVASI, CHEYYAR T.K -
Veerampakkam
TAMILNADU 604408

Coordinator Contact Details:
Email: clinq@cmcvellore.ac.in
Contact Number: 0416-2283102

E. Rao
Medical Officer
Veerambakkam Primary
Health Centre
Cheyyar HUD Pin - 604 408

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential

https://www.cmcvellore.ac.in/clinq/ViewMonthlySummary/printResult_new.aspx?resDetId=268&noSamples=1&month=8&year=2023&groupId=9&yearGroup=110