

9/28/2023

External Quality Assurance Scheme - Print Monthly Summary



## CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name Govt. PHC, Veerampakkam  
Constituent Group Chemistry III  
PT item Lyophilized human serum based

Lab No 9264  
Date of Result Entered : 19/07/2023  
Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	108.24	4.71	5.10	108 mg/dL	-0.05	1.70
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	5.84	5.91	0.34	5.6 mg/dL	-0.70	0.12
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	113.73	3.56	4.04	121 mg/dL	1.80	1.39

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PHC, Veerampakkam  
THE MEDICAL OFFICER , GOVT. PRIMARY HEALTH CENTER, VEERAMPAKKAM, VANDAVASI, CHEYYAR T.K -  
Veerampakkam  
TAMILNADU604408

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*S. Senthil*  
**Medical Officer**  
Veerambakkam Primary  
Health Centre  
Cheyyar HUD Pin : 604 408

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential