

9/28/2023

External Quality Assurance Scheme - Print Monthly Summary



## CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name Govt. PHC, Veerampakkam  
Constituent Group Chemistry III  
PT item Lyophilized human serum based

Lab No 9264  
Date of Result Entered : 19/06/2023  
Date of Report Published : 04/07/2023

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name | No of Participants | DV     | Participants |       | Your Value | SDI   | U    |
|-------|----------------|---------------------------------------|---------------|--------------------|--------|--------------|-------|------------|-------|------|
|       |                |                                       |               |                    |        | CV           | SD    |            |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Biosystems    | 36                 | 280.51 | 4.07         | 11.42 | 274 mg/dL  | -0.57 | 3.81 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Biosystems    | 39                 | 2.67   | 6.50         | 0.17  | 2.6 mg/dL  | -0.40 | 0.06 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Biosystems    | 37                 | 113.79 | 4.23         | 4.81  | 118 mg/dL  | 0.88  | 1.58 |

| SDI Range             | Interpretation                           |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent.                               |
| Within ±1.01 to ±2.00 | Good.                                    |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal.     |
| Beyond ±3.0           | Unacceptable performance. Action Signal. |

LAB ADDRESS :  
Govt. PHC, Veerampakkam  
THE MEDICAL OFFICER , GOVT. PRIMARY HEALTH CENTER,VEERAMPAKKAM, VANDAVASI,CHEYAR T.K -  
Veerampakkam  
TAMILNADU604408

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*S. S. S.*  
Medical Officer  
Veerambakkam Primary  
Health Centre  
Cheyyar HUD Pin : 604 408

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential