

9/28/2023

External Quality Assurance Scheme - Print Monthly Summary



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name Govt. PHC, Veerampakkam
Constituent Group Chemistry III
PT item Lyophilized human serum based

Lab No 9264
Date of Result Entered : 20/04/2023
Date of Report Published : 06/05/2023

Item No	Test Name	Method	Instrument	Number of Samples	Mean	SD	CV (%)	Reference Range	SDI	SDI Range
1	GLUCOSE III	GOD-POD III	Biosystems	35	118.67	7.30	8.67	130 mg/dL	1.31	2.93
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15	109 mg/dL	712.13	0.05
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03	2.3 mg/dL	-18.32	2.07

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Veerampakkam
THE MEDICAL OFFICER, GOVT. PRIMARY HEALTH CENTER, VEERAMPAKKAM, VANDAVASI, CHEYYAR T.K -
Veerampakkam
TAMILNADU 604408

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

E. S. S. S.
Medical Officer
Veerampakkam Primary
Health Centre
Dist: Vellore Pin: 604 408

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential

https://www.cmcvellore.ac.in/clinqc/ViewMonthlySummary/printResult_new.aspx?resDetId=264&noSamples=1&month=4&year=2023&groupId=9&yearGroup=110