

12:32:09

NAME: CHOLE

mg Stored Value

Res: 0.040

Factor: 500.000

No: 1

Res: 0.223

Res: 111.4593mg/dl LO

Delta: 0.250

Res: 5.7748 HI

St. No: 3

Res: 0.461

Res: 106.9965mg/dl NM

St. No: 4

Res: 0.309

Res: 71.7274mg/dl NM

# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023

Lab No

12207

Date of Result Entered :

17/07/2023

Date of Report Published :

07/08/2023

Govt. PHC, Ganapathipalayam

Chemistry III

Unphylized human serum based

Method / Principle Name

Analyzer Name

No of Participants

DV

Participants CV

SD

Your Value

SDI

U

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	106.99 mg/dL	-0.39	0.28
Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.77 mg/dL	-0.17	0.02
CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	111.45 mg/dL	-0.33	0.22
Interpretation								
SDI Range		Excellent.						
Within -1.00 to +1.00		Good.						
Within ±1.01 to ±2.00		Accept with caution. Warning Signal.						
Within ±2.01 to ±2.99		Unacceptable performance. Action Signal.						
Beyond ±3.0								

LAB ADDRESS :  
 Govt. PHC, Ganapathipalayam  
 THE MEDICAL OFFICER, MODAKURICHI  
 Ganapathipalayam  
 TAMILNADU638152


Coordinator Contact Details:  
 Email: cingc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christodoss*

Dr. Pamela Christodoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

*N. Anbar*

மருத்துவ அலுவலர்  
 அரசு ஆரம்ப சுகாதார நிலையம்  
 கலைத்தியானையம்



PC-1024