

07/06/23 12:12:18

TEST NAME: CREAT

Using Stored Value

Std Res: 0.087

Factor: 23.108

SI No: 1

Init.OD: 0.413 HI

Delta: 0.103

Res: 2.3816 HI

ABS: 0.304

Res: 109.2375mg/dl LO

ABS: 1.107

Res: 285.1249mg/dl HI

External Quality Assurance Scheme - Print Monthly Summary

# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023

June-23



t. PHC, Ganapathipalayam

mistry III

phlized human serum based

Lab No

12207

Date of Result Entered :

20/06/2023

Date of Report Published :

04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	285.12 mg/dL	-0.10	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.38 mg/dL	-1.11	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	109.23 mg/dL	-0.89	0.24
Interpretation										
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

LAB ADDRESS :

Govt. PHC, Ganapathipalayam

THE MEDICAL OFFICER, MODAKURICHI

Ganapathipalayam

TAMILNADU638152

Coordinator Contact Details:

Email: cilingc@cmcvellore.ac.in

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*N. Anbar*

மருத்துவ அலுவலர்  
அரசு ஆரம்ப சுகாதார நிலையம்  
கண்ணாடிபாளையம்

PC-102