

9/28/2023

External Quality Assurance Scheme - Print Monthly Summary



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. PHC, Veerampakkam
Constituent Group Chemistry III
PT item Lyophilized human serum based

Lab No 9264
Date of Result Entered : 18/03/2023
Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	92.67	8.46	7.84	91 mg/dL	-0.21	2.61
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.37	9.41	0.13	1.2 mg/dL	-1.32	0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	112.13	5.96	6.68	115 mg/dL	0.43	2.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Veerampakkam
THE MEDICAL OFFICER, GOVT. PRIMARY HEALTH CENTER, VEERAMPAKKAM, VANDAVASI, CHEYYAR T.K -
Veerampakkam
TAMILNADU 604408

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

G. S. S.
Medical Officer
Veerambakkam Primary
Health Centre
Cheyyar HUD Pin : 604 4008

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential