

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



May 23

20/05/23
12:06:02

TEST NAME: GLUCOSE
Using Stored Value
Blk Res: 0.020
Using Stored Value
Std Res: 0.414
Factor: 241.654
Sr. No: 1
ABS: 0.763
Res: 184.4130mg/dl HI
Sr. No: 9
Init. OD: 0.468 HI
Delta: 0.079
Res: 1.7913 HI

Govt. PHC, Ganapathipalayam
Chemistry III
Polypurified human serum based

Lab No: 12207
Date of Result Entered: 20/05/2023
Date of Report Published: 05/06/2023

SL.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic)	1834	180.88	5.19	9.39	184.41 mg/dL	0.38	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic)	1854	1.76	7.48	0.13	1.79 mg/dL	0.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic) / Fully Automatic)	1798	114.60	5.08	5.83	106.77 mg/dL	-1.34	0.27
Interpretation										
SDI Range										
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

LAB ADDRESS :
Govt. PHC, Ganapathipalayam
THE MEDICAL OFFICER, MODAKURICHI
Ganapathipalayam
TAMILNADU638152

Coordinator Contact Details:
Email: cllngc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

N. Anand
மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்
கரையூதிபாளையம்