



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name: Govt. UGPHC, Muthupettai
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 7269
Date of Result Entered: 20/04/2023
Date of Report Published: 06/05/2023

APRIL-23

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	1-0			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	119 mg/dL	-0.06	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.2 mg/dL	0.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	117 mg/dL	0.93	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UGPHC, Muthupettai
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVARUR DISTRICT.
Muthupettai
MILNADU614704

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

MEDICAL OFFICER
Govt. Primary Health Centre
MUTHUPET - 614 704
Thiruvavur District