

CALIBRATION CERTIFICATE												
CERTIFICATE NO: SBS/CL/23/11755					MEDICAL DEVICES					Page	No:1 of 1	
Issue Date				26-09-2023								
SRF No & Date					SRF/23/00405-0002 & 25-09-2023							
Receipt Date					25-09-2023							
Calibration Date					25-09-2023							
Calibration I	Due		-		24-09	9-2024						
Customer Name & Address												
GOVERNMENT UPGRADED PRIMARY HEALTH CENTRE,												
PONGALUR-641667, TIRUPPUR DISTRICT.												
Details of Device Under Calibration (DUC)												
Description : ELECTRICAL SAFE (MICROSCOPE)		ſΥ	Make	& Model		:		LABOMED & LX-400				
Range :		MULTI	MULTI			Sr. No		:		1583339		
Resolution : MULTI				Identi	fication No		:		NA			
DUC Conditio	n :	SATIS	FACTORY		Locat	ion		:		LAB		
Environmental Conditions & Calibration Procedure Details												
Environmental Details Temperature			25.6°C Relative Hum		nidit	nidity		52% RH				
Calibration Procedure No SBS/CP/MD/2				29 Calibration done at ONSITE			ONSITE					
				Refere	nce S	tandards Det	tails	5				
S.No Description				Make/ SI No:			Certificate N		ate No	D		Validity
1 Electrical Safety Analyser				Rigel Medical & 44L-1059		M	M-230809-16-4		-4		10-08-2024	
ELECTRICAL SAFETY												
RESULTS			-									
S S S S S S S S S S S S S S S S S S S				MEASURED VALUES								

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	97	13.92
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)
	<5000µAfor B,BF,CF	125	8.0
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)
	<500µAfor B,BF,CF	225	7.2

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

2. The user should determine the suitability of the instrument for its intended use.

3. The recalibration interval should be determined on the user requirement.

4. The results stated in this certificate relate only to the item calibrated.

5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.

6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

(Calibration Engineer) . K.SATHYAMOORTHY

Calibrated By

/	MEDICA
EB	Chennai
VIHS	600 032 UT
12	5 * 5

Technical Manager C.SHANMUGARAJ

Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.