



# BLUEMED HEALTHCARE

*dedication beyond measure*



## CALIBRATION PERFORMANCE FOR CENTRIFUGE

Calibration Certificate No: CAL/GPHC-VBM/2023/003

Customer Name and Address	M/s.Govt. Primary Health Centre Veerambakkam, Cheyyar Taluk, Tamilnadu – 604408.	Equipment ID :	*****
Department :	LABORATORY	Serial No :	JE2409L1
Make :	NA	Date of Calibration :	27-03-2023
Model :	NA	Next Calibration Due:	27-03-2024

### Traceability Details:

S.No	Standards Used	Make	Model	Serial No	Calibrated on	Calibration Due on
1	Tacho Meter	Lutron	DT-2234C	L140351	15-02-2023	14-02-2024
2	Electrical Safety Analyzer	RIGEL MEDICAL	RIGEL 288+	05H-0600	09-08-2022	08-08-2024

Equipment Condition at the time of Calibration: Functioning Good

### Calibration Results:

S.No	Test Parameters	Unit	Set value	Measured value	Acceptable Range	Remarks
1	Rotation Per Minute	RPM	1000	990	±5%	Pass
			2000	1988	±5%	Pass
			3000	2987	±5%	Pass

Electrical safety Test Passed (Report Attached)

### Remarks:

1. The Standards used are traceable to National/International Standards.
2. The Result reported is valid at the time of and under the stated conditions of measurement.
3. This report refers only to the item calibrated and shall not be reproduced in full without the written approval of the issuing laboratory.
4. The above equipment calibrated at site, with TEMP/RH Level: 25±10°C/40–70%RH.
5. The above Tolerance level has been defined from the manufacturer's manual.

Calibrated by:

Calibration Engineer



Approved by:

Authorized Signatory

