

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Sekodi

[Details About Robust Analysis](#)

Lab No 12119

[Detail About Monthly Summary](#)

Month May

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	L
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	69	181.20	5.81	10.52	180 mg/dL	-0.11	2.53
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	68	1.75	7.05	0.12	1.7 mg/dL	-0.41	0.03
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	154	114.84	5.08	5.83	111 mg/dL	-0.66	0.94

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

MEDICAL OFFICER
GOVT PRIMARY HEALTH CENTRE
SEKKODI
PALACODE BLOCK
District: Kolar

Links