

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Sekodi  
 Lab No 12119  
 Month April  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Year Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation )	74	118.33	5.77	6.83	127 mg/dL	1.27	1.59
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation )	65	2.14	7.02	0.15	2 mg/dL	-0.93	0.04
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	143	111.22	4.31	4.79	103 mg/dL	-1.72	0.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

  
 MEDICAL OFFICER  
 GOVT. PRIMARY HEALTH CENTRE  
 SEKKODI,  
 PALACODE, DIST.