



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 12.09.2023
	Calibration Valid: 11.03.2024
I.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, PRIMARY HEALRH CENTRE, SOUTH VIJAYA NARAYANAM, THIRUNELVELI.
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	CENTRIFUGE
i Manufactured by	LABOTECH
ii Model	BD INSTRUMENTATION
iii Serial No	BDI 152
iv Location	LABORATORY

CALIBRATION WORKSHEET

Tabulation For – Centrifuge


S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	1000	0
2	1200	1208	8
3	1500	1505	5
4	2000	1997	-3

Acceptable deviation is ± 5 RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016
GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106
88833 15662

No.42/1, Navaindia Road, Peelamec
CBE – 04, hbmedicalsystem@gmail.c



S.no	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	05.0	05.1	0.1
2	10.0	10.2	0.2
3	15.0	15.1	0.1
4	20.0	20.2	0.2

Acceptance deviation: ± 0.5 seconds

Testy equipment used

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer (non - contact) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

Remarks

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by :