



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023

AUGUST-2023



PC-1024

Lab Name: Govt. PHC, Thulukkamuthur

Lab No: 9270

Institution Group: Chemistry III

Date of Result Entered: 19/08/2023

Item: Lyophilized human serum based

Date of Report Published: 01/09/2023

No	Analyte	Method / Principle Name	Equipment Name	Mean	SD	Participants	Your Value	SDI	U	
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	171.7 mg/dL	-0.43	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	2.01 mg/dL	2.10	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	114.4 mg/dL	0.13	0.22

SDI Range	Interpretation
within -1.00 to +1.00	Excellent.
within ±1.01 to ±2.00	Good.
within ±2.01 to ±2.99	Accept with caution. Warning Signal.
beyond ±3.0	Unacceptable performance. Action Signal.

**ADDRESS :**

Govt. PHC, Thulukkamuthur  
MEDICAL OFFICER, GOVERNMENT HEALTH PRIMARY CENTRE, THULUKKAMUTHUR, THULUKKAMUTHUR (PO), AVINASHI (TK),  
TAMILNADU (DT)  
Thulukkamuthur  
TAMILNADU 641654

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*MT*  
MEDICAL OFFICER  
GOVT. PRAMARY HEALTH CENTRE  
THULUKKAMUTHUR  
AVINASHI - 641 654