

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Gangaikondan

[Details About Robust Analysis](#)

Lab No 12071

[Detail About Monthly Summary](#)

Month February

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)


[Print](#)

Date of Result Entered : 20/02/2023

Date of Report Published : 08/03/2023

1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	155.11	6.56	10.18	149.9 mg/dL	-0.51	0.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1872	3.40	9.00	0.31	3.2 mg/dL	-0.65	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1788	112.79	5.23	5.90	108 mg/dL	-0.81	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

[Links](#)


MEDICAL OFFICER
 Govt. Primary Health Centre
 Gangaikondan