



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023

JULY-2023



Name: Govt. PHC, Thulukkamuthur Lab No: 9270
 Constituent Group: Chemistry III Date of Result Entered: 18/07/2023
 Item: Lyophilized human serum based Date of Report Published: 07/08/2023

No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	C
						CV	SD			
	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	104.7 mg/dL	-0.77	0.28
	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.71 mg/dL	2.51	0.02
	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	119 mg/dL	1.29	0.22

SDI Range	Interpretation
within -1.00 to +1.00	Excellent.
within ±1.01 to ±2.00	Good.
within ±2.01 to ±2.99	Accept with caution. Warning Signal.
around ±3.0	Unacceptable performance. Action Signal.

3 ADDRESS :

Govt. PHC, Thulukkamuthur
 Medical Officer, Government Health Primary Centre, Thulukkamuthur, Thulukkamuthur (PO), Avinashi (TK),
 Uppur (DT)
 Thulukkamuthur
 AVINASHI-641654

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

my
 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 THULUKKAMUTHUR
 AVINASHI - 641 654